

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

BC1709B

2017-11-10



Instrument	Parameter	Low		Normal		High		+
		LOT	BC1709BL	LOT	BC1709BN	LOT	BC1709BH	
BC-5800,BC-5600 QC Mode	WBC $\times 10^9/L$	3.40	$\pm 0.50$	7.85	$\pm 1.00$	17.60	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.68	$\pm 0.35$	4.40	$\pm 0.65$	11.53	$\pm 1.45$	
	Lym# $\times 10^9/L$	1.28	$\pm 0.35$	2.28	$\pm 0.65$	3.26	$\pm 1.45$	
	Mon# $\times 10^9/L$	0.26	$\pm 0.25$	0.59	$\pm 0.55$	1.32	$\pm 1.25$	
	Eos# $\times 10^9/L$	0.15	$\pm 0.15$	0.51	$\pm 0.45$	1.32	$\pm 1.25$	
	Bas# $\times 10^9/L$	0.03	$\pm 0.03$	0.08	$\pm 0.08$	0.18	$\pm 0.18$	
	Neu%	49.5	$\pm 9.0$	56.0	$\pm 8.0$	65.5	$\pm 8.0$	
	Lym%	37.5	$\pm 9.0$	29.0	$\pm 8.0$	18.5	$\pm 8.0$	
	Mon%	7.5	$\pm 6.5$	7.5	$\pm 6.5$	7.5	$\pm 7.0$	
	Eos%	4.5	$\pm 4.0$	6.5	$\pm 5.5$	7.5	$\pm 7.0$	
	Bas%	1.0	$\pm 1.0$	1.0	$\pm 1.0$	1.0	$\pm 1.0$	
	RBC $\times 10^{12}/L$	2.06	$\pm 0.18$	4.44	$\pm 0.24$	5.07	$\pm 0.30$	
	HGB g/L	59	$\pm 4$	138	$\pm 6$	167	$\pm 8$	
	HCT %	17.9	$\pm 1.5$	42.0	$\pm 2.0$	52.2	$\pm 2.4$	
	MCV fL	87.0	$\pm 5.0$	94.5	$\pm 5.0$	103.0	$\pm 5.0$	
	MCH pg	28.6	$\pm 2.5$	31.1	$\pm 2.5$	32.9	$\pm 2.5$	
	MCHC g/L	329	$\pm 30$	329	$\pm 30$	320	$\pm 30$	
	RDW-CV %	14.0	$\pm 3.0$	14.2	$\pm 3.0$	13.5	$\pm 3.0$	
	RDW-SD fL	42.0	$\pm 8.0$	45.0	$\pm 8.0$	48.0	$\pm 8.0$	
	PLT $\times 10^9/L$	53	$\pm 20$	242	$\pm 40$	487	$\pm 60$	
	MPV fL	7.6	$\pm 3.0$	8.3	$\pm 3.0$	9.0	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.200	$\pm 0.100$	0.445	$\pm 0.200$	
	PDW*	18.1	$\pm 3.0$	18.2	$\pm 3.0$	16.2	$\pm 3.0$	
	P-LCC $\times 10^9/L$	15	$\pm 15$	74	$\pm 25$	135	$\pm 35$	
	P-LCR %	24.5	$\pm 10.0$	30.0	$\pm 10.0$	28.0	$\pm 10.0$	
BC-5500,BC-5200 QC Mode	WBC $\times 10^9/L$	3.35	$\pm 0.50$	7.55	$\pm 1.00$	17.05	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.71	$\pm 0.35$	4.42	$\pm 0.65$	11.17	$\pm 1.40$	
	Lym# $\times 10^9/L$	1.26	$\pm 0.35$	2.15	$\pm 0.65$	3.15	$\pm 1.40$	
	Mon# $\times 10^9/L$	0.20	$\pm 0.20$	0.34	$\pm 0.34$	1.19	$\pm 1.05$	
	Eos# $\times 10^9/L$	0.15	$\pm 0.15$	0.57	$\pm 0.50$	1.36	$\pm 1.20$	
	Bas# $\times 10^9/L$	0.03	$\pm 0.03$	0.08	$\pm 0.08$	0.17	$\pm 0.17$	
	Neu%	51.0	$\pm 9.0$	58.5	$\pm 8.0$	65.5	$\pm 8.0$	
	Lym%	37.5	$\pm 9.0$	28.5	$\pm 8.0$	18.5	$\pm 8.0$	
	Mon%	6.0	$\pm 5.5$	4.5	$\pm 4.0$	7.0	$\pm 6.0$	
	Eos%	4.5	$\pm 4.0$	7.5	$\pm 6.0$	8.0	$\pm 7.0$	
	Bas%	1.0	$\pm 1.0$	1.0	$\pm 1.0$	1.0	$\pm 1.0$	
	RBC $\times 10^{12}/L$	2.08	$\pm 0.18$	4.50	$\pm 0.24$	5.09	$\pm 0.30$	
	HGB g/L	63	$\pm 4$	147	$\pm 6$	178	$\pm 8$	
	HCT %	17.3	$\pm 1.5$	40.5	$\pm 2.0$	49.6	$\pm 2.4$	
	MCV fL	83.0	$\pm 5.0$	90.0	$\pm 5.0$	97.5	$\pm 5.0$	
	MCH pg	30.3	$\pm 2.5$	32.7	$\pm 2.5$	35.0	$\pm 2.5$	
	MCHC g/L	365	$\pm 30$	363	$\pm 30$	359	$\pm 30$	
	RDW-CV %	11.0	$\pm 3.0$	11.5	$\pm 3.0$	11.0	$\pm 3.0$	
	RDW-SD fL	28.5	$\pm 8.0$	33.5	$\pm 8.0$	37.0	$\pm 8.0$	
	PLT $\times 10^9/L$	51	$\pm 20$	235	$\pm 40$	472	$\pm 60$	
	MPV fL	8.1	$\pm 3.0$	9.0	$\pm 3.0$	9.8	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.210	$\pm 0.100$	0.460	$\pm 0.200$	
	PDW*	18.2	$\pm 3.0$	18.2	$\pm 3.0$	16.5	$\pm 3.0$	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES


**BC1709B**  
**2017-11-10**

Instrument	Parameter	Low		Normal		High		++
		LOT	BC1709BL	LOT	BC1709BN	LOT	BC1709BH	
<b>BC-5390</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.20	$\pm 0.50$	7.45	$\pm 1.00$	16.75	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.72	$\pm 0.30$	4.44	$\pm 0.60$	11.39	$\pm 1.55$	
	Lym# $\times 10^9/L$	1.16	$\pm 0.30$	2.05	$\pm 0.60$	2.98	$\pm 1.20$	
	Mon# $\times 10^9/L$	0.18	$\pm 0.15$	0.42	$\pm 0.35$	1.07	$\pm 0.85$	
	Eos# $\times 10^9/L$	0.15	$\pm 0.15$	0.54	$\pm 0.45$	1.31	$\pm 1.05$	
	Bas# $\times 10^9/L$	0.79	$\pm 0.35$	2.03	$\pm 0.75$	5.19	$\pm 1.70$	
	Neu%	53.6	$\pm 9.0$	59.6	$\pm 8.0$	68.0	$\pm 9.0$	
	Lym%	36.2	$\pm 9.0$	27.5	$\pm 8.0$	17.8	$\pm 7.0$	
	Mon%	5.6	$\pm 4.0$	5.6	$\pm 4.0$	6.4	$\pm 5.0$	
	Eos%	4.6	$\pm 4.0$	7.3	$\pm 6.0$	7.8	$\pm 6.0$	
	Bas%	24.6	$\pm 10.0$	27.3	$\pm 10.0$	31.0	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.01	$\pm 0.18$	4.43	$\pm 0.24$	5.05	$\pm 0.30$	
	HGB g/L	54	$\pm 4$	128	$\pm 6$	156	$\pm 8$	
	HCT %	17.1	$\pm 1.5$	40.8	$\pm 2.0$	50.2	$\pm 2.4$	
	MCV fL	85.0	$\pm 5.0$	92.0	$\pm 5.0$	99.5	$\pm 5.0$	
	MCH pg	26.9	$\pm 2.5$	28.9	$\pm 2.5$	30.9	$\pm 2.5$	
	MCHC g/L	316	$\pm 30$	314	$\pm 30$	310	$\pm 30$	
	RDW-CV %	13.5	$\pm 3.0$	14.3	$\pm 3.0$	13.8	$\pm 3.0$	
	RDW-SD fL	44.0	$\pm 8.0$	47.0	$\pm 8.0$	51.0	$\pm 8.0$	
	PLT $\times 10^9/L$	47	$\pm 20$	218	$\pm 40$	486	$\pm 60$	
	MPV fL	11.0	$\pm 3.0$	11.6	$\pm 3.0$	11.6	$\pm 3.0$	
<b>BC-5390 CRP</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.25	$\pm 0.50$	7.55	$\pm 1.00$	17.20	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.72	$\pm 0.30$	4.42	$\pm 0.65$	11.61	$\pm 1.55$	
	Lym# $\times 10^9/L$	1.15	$\pm 0.30$	2.08	$\pm 0.65$	3.10	$\pm 1.25$	
	Mon# $\times 10^9/L$	0.23	$\pm 0.20$	0.49	$\pm 0.40$	1.03	$\pm 0.90$	
	Eos# $\times 10^9/L$	0.15	$\pm 0.15$	0.57	$\pm 0.50$	1.46	$\pm 1.25$	
	Bas# $\times 10^9/L$	0.80	$\pm 0.35$	2.06	$\pm 0.80$	5.35	$\pm 1.75$	
	Neu%	53.0	$\pm 9.0$	58.5	$\pm 8.0$	67.5	$\pm 9.0$	
	Lym%	35.5	$\pm 9.0$	27.5	$\pm 8.0$	18.0	$\pm 7.0$	
	Mon%	7.0	$\pm 5.0$	6.5	$\pm 5.0$	6.0	$\pm 5.0$	
	Eos%	4.5	$\pm 4.0$	7.5	$\pm 6.0$	8.5	$\pm 7.0$	
	Bas%	24.7	$\pm 10.0$	27.3	$\pm 10.0$	31.1	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.04	$\pm 0.18$	4.43	$\pm 0.24$	5.08	$\pm 0.30$	
	HGB g/L	56	$\pm 4$	130	$\pm 6$	157	$\pm 8$	
	HCT %	17.0	$\pm 1.5$	40.8	$\pm 2.0$	50.8	$\pm 2.4$	
	MCV fL	83.5	$\pm 5.0$	92.0	$\pm 5.0$	100.0	$\pm 5.0$	
	MCH pg	27.5	$\pm 2.5$	29.3	$\pm 2.5$	30.9	$\pm 2.5$	
	MCHC g/L	329	$\pm 30$	319	$\pm 30$	309	$\pm 30$	
	RDW-CV %	13.5	$\pm 3.0$	14.3	$\pm 3.0$	14.0	$\pm 3.0$	
	RDW-SD fL	41.0	$\pm 8.0$	47.0	$\pm 8.0$	49.5	$\pm 8.0$	
	PLT $\times 10^9/L$	43	$\pm 20$	208	$\pm 40$	471	$\pm 60$	
	MPV fL	8.8	$\pm 3.0$	9.1	$\pm 3.0$	9.1	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.185	$\pm 0.100$	0.425	$\pm 0.200$	
	PDW*	17.0	$\pm 3.0$	17.1	$\pm 3.0$	16.2	$\pm 3.0$	
	P-LCC $\times 10^9/L$	15	$\pm 15$	56	$\pm 25$	105	$\pm 35$	
	P-LCR %	25.0	$\pm 10.0$	27.5	$\pm 10.0$	22.5	$\pm 10.0$	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

BC1709B



2017-11-10

Instrument	Parameter	Low		Normal		High		+++
		LOT	BC1709BL	LOT	BC1709BN	LOT	BC1709BH	
BC-5300,BC-5100	WBC $\times 10^9/L$		3.25 $\pm$ 0.50		7.65 $\pm$ 1.00		17.25 $\pm$ 2.50	
BC-5380,BC-5180	Neu# $\times 10^9/L$		1.80 $\pm$ 0.30		4.72 $\pm$ 0.65		11.76 $\pm$ 1.40	
QC Mode	Lym# $\times 10^9/L$		1.16 $\pm$ 0.30		2.07 $\pm$ 0.65		3.02 $\pm$ 1.40	
(Software version lower than 1.24.00.16860)	Mon# $\times 10^9/L$		0.12 $\pm$ 0.12		0.29 $\pm$ 0.29		0.95 $\pm$ 0.80	
	Eos# $\times 10^9/L$		0.16 $\pm$ 0.15		0.57 $\pm$ 0.50		1.52 $\pm$ 1.25	
	Bas# $\times 10^9/L$		1.76 $\pm$ 0.35		4.99 $\pm$ 0.80		13.37 $\pm$ 1.75	
	Neu%		55.5 $\pm$ 9.0		61.7 $\pm$ 8.0		68.2 $\pm$ 8.0	
	Lym%		35.7 $\pm$ 9.0		27.0 $\pm$ 8.0		17.5 $\pm$ 8.0	
	Mon%		3.8 $\pm$ 3.8		3.8 $\pm$ 3.8		5.5 $\pm$ 4.5	
	Eos%		5.0 $\pm$ 4.0		7.5 $\pm$ 6.0		8.8 $\pm$ 7.0	
	Bas%		54.2 $\pm$ 10.0		65.2 $\pm$ 10.0		77.5 $\pm$ 10.0	
	RBC $\times 10^{12}/L$		2.04 $\pm$ 0.18		4.44 $\pm$ 0.24		5.05 $\pm$ 0.30	
	HGB g/L		56 $\pm$ 4		130 $\pm$ 6		157 $\pm$ 8	
	HCT %		18.4 $\pm$ 1.5		43.7 $\pm$ 2.0		53.4 $\pm$ 2.4	
	MCV fL		90.0 $\pm$ 5.0		98.5 $\pm$ 5.0		105.8 $\pm$ 5.0	
	MCH pg		27.5 $\pm$ 2.5		29.3 $\pm$ 2.5		31.1 $\pm$ 2.5	
	MCHC g/L		305 $\pm$ 30		297 $\pm$ 30		294 $\pm$ 30	
	RDW-CV %		13.8 $\pm$ 3.0		14.2 $\pm$ 3.0		13.5 $\pm$ 3.0	
	RDW-SD fL		53.8 $\pm$ 8.0		59.7 $\pm$ 8.0		63.0 $\pm$ 8.0	
	PLT $\times 10^9/L$		46 $\pm$ 20		212 $\pm$ 40		462 $\pm$ 60	
	MPV fL		8.4 $\pm$ 3.0		9.0 $\pm$ 3.0		9.1 $\pm$ 3.0	
	PCT %*		0.050 $\pm$ 0.050		0.188 $\pm$ 0.100		0.418 $\pm$ 0.200	
	PDW*		17.7 $\pm$ 3.0		17.6 $\pm$ 3.0		16.2 $\pm$ 3.0	
BC-5300,BC-5100	WBC $\times 10^9/L$		3.25 $\pm$ 0.50		7.50 $\pm$ 1.00		17.00 $\pm$ 2.50	
BC-5380,BC-5180	Neu# $\times 10^9/L$		1.79 $\pm$ 0.30		4.63 $\pm$ 0.65		11.75 $\pm$ 1.40	
QC Mode	Lym# $\times 10^9/L$		1.18 $\pm$ 0.30		2.08 $\pm$ 0.65		2.94 $\pm$ 1.40	
(Software version 1.24.00.16860 or higher)	Mon# $\times 10^9/L$		0.13 $\pm$ 0.13		0.23 $\pm$ 0.23		0.88 $\pm$ 0.70	
	Eos# $\times 10^9/L$		0.16 $\pm$ 0.15		0.56 $\pm$ 0.45		1.43 $\pm$ 1.20	
	Bas# $\times 10^9/L$		1.84 $\pm$ 0.35		5.00 $\pm$ 0.80		13.29 $\pm$ 1.70	
	Neu%		55.0 $\pm$ 9.0		61.7 $\pm$ 8.0		69.1 $\pm$ 8.0	
	Lym%		36.2 $\pm$ 9.0		27.7 $\pm$ 8.0		17.3 $\pm$ 8.0	
	Mon%		4.0 $\pm$ 3.5		3.1 $\pm$ 3.1		5.2 $\pm$ 4.0	
	Eos%		4.8 $\pm$ 4.0		7.5 $\pm$ 6.0		8.4 $\pm$ 7.0	
	Bas%		56.6 $\pm$ 10.0		66.6 $\pm$ 10.0		78.2 $\pm$ 10.0	
	RBC $\times 10^{12}/L$		2.03 $\pm$ 0.18		4.44 $\pm$ 0.24		5.06 $\pm$ 0.30	
	HGB g/L		56 $\pm$ 4		130 $\pm$ 6		158 $\pm$ 8	
	HCT %		17.8 $\pm$ 1.5		42.2 $\pm$ 2.0		52.1 $\pm$ 2.4	
	MCV fL		87.5 $\pm$ 5.0		95.0 $\pm$ 5.0		103.0 $\pm$ 5.0	
	MCH pg		27.6 $\pm$ 2.5		29.3 $\pm$ 2.5		31.2 $\pm$ 2.5	
	MCHC g/L		315 $\pm$ 30		308 $\pm$ 30		303 $\pm$ 30	
	RDW-CV %		13.5 $\pm$ 3.0		14.0 $\pm$ 3.0		13.5 $\pm$ 3.0	
	RDW-SD fL		49.5 $\pm$ 8.0		55.5 $\pm$ 8.0		59.0 $\pm$ 8.0	
	PLT $\times 10^9/L$		44 $\pm$ 20		210 $\pm$ 40		471 $\pm$ 60	
	MPV fL		8.9 $\pm$ 3.0		9.3 $\pm$ 3.0		9.4 $\pm$ 3.0	
	PCT %*		0.050 $\pm$ 0.050		0.190 $\pm$ 0.100		0.440 $\pm$ 0.200	
	PDW*		16.9 $\pm$ 3.0		17.0 $\pm$ 3.0		16.2 $\pm$ 3.0	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

**CONTROL**

ASSAY VALUES AND EXPECTED RANGES

**LOT****BC1709B****2017-11-10**

Instrument	Parameter	Low		Normal		High		++++
		LOT	BC1709BL	LOT	BC1709BN	LOT	BC1709BH	
<b>BC-5000,BC-5150,BC-5120</b> <b>BC-5130,BC-5140,BC-5000VET</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.30	$\pm 0.50$	7.70	$\pm 1.00$	17.00	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.72	$\pm 0.40$	4.44	$\pm 0.95$	11.07	$\pm 2.05$	
	Lym# $\times 10^9/L$	1.17	$\pm 0.30$	2.06	$\pm 0.65$	2.89	$\pm 1.20$	
	Mon# $\times 10^9/L$	0.24	$\pm 0.24$	0.57	$\pm 0.57$	1.50	$\pm 1.50$	
	Eos# $\times 10^9/L$	0.15	$\pm 0.15$	0.55	$\pm 0.55$	1.33	$\pm 1.33$	
	Bas# $\times 10^9/L$	0.02	$\pm 0.02$	0.09	$\pm 0.09$	0.22	$\pm 0.22$	
	Neu%	52.0	$\pm 12.0$	57.6	$\pm 12.0$	65.1	$\pm 12.0$	
	Lym%	35.4	$\pm 9.0$	26.7	$\pm 8.0$	17.0	$\pm 7.0$	
	Mon%	7.4	$\pm 7.4$	7.4	$\pm 7.4$	8.8	$\pm 8.8$	
	Eos%	4.5	$\pm 4.5$	7.1	$\pm 7.1$	7.8	$\pm 7.8$	
	Bas%	0.7	$\pm 0.7$	1.2	$\pm 1.2$	1.3	$\pm 1.3$	
	RBC $\times 10^{12}/L$	2.02	$\pm 0.18$	4.45	$\pm 0.24$	5.06	$\pm 0.30$	
	HGB g/L	55	$\pm 4$	132	$\pm 6$	160	$\pm 8$	
	HCT %	17.9	$\pm 1.5$	42.1	$\pm 2.0$	50.9	$\pm 2.4$	
	MCV fL	88.5	$\pm 5.0$	94.5	$\pm 5.0$	100.5	$\pm 5.0$	
	MCH pg	27.2	$\pm 2.5$	29.7	$\pm 2.5$	31.6	$\pm 2.5$	
	MCHC g/L	308	$\pm 30$	314	$\pm 30$	315	$\pm 30$	
	RDW-CV %	16.0	$\pm 3.0$	17.0	$\pm 3.0$	16.0	$\pm 3.0$	
	RDW-SD fL	51.0	$\pm 8.0$	58.5	$\pm 8.0$	62.5	$\pm 8.0$	
	PLT $\times 10^9/L$	51	$\pm 20$	237	$\pm 40$	485	$\pm 60$	
	MPV fL	10.0	$\pm 3.0$	10.2	$\pm 3.0$	10.8	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.245	$\pm 0.100$	0.520	$\pm 0.200$	
	PDW*	17.2	$\pm 3.0$	17.5	$\pm 3.0$	16.4	$\pm 3.0$	
	P-LCC $\times 10^9/L$ **	15	$\pm 15$	80	$\pm 25$	145	$\pm 35$	
	P-LCR %**	32.0	$\pm 10.0$	33.5	$\pm 10.0$	31.0	$\pm 10.0$	
<b>BC-5300Vet,BC-5100Vet</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.25	$\pm 0.50$	7.65	$\pm 1.00$	17.25	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.80	$\pm 0.30$	4.72	$\pm 0.65$	11.76	$\pm 1.40$	
	Lym# $\times 10^9/L$	1.16	$\pm 0.30$	2.07	$\pm 0.65$	3.02	$\pm 1.40$	
	Mon# $\times 10^9/L$	0.12	$\pm 0.12$	0.29	$\pm 0.29$	0.95	$\pm 0.80$	
	Eos# $\times 10^9/L$	0.16	$\pm 0.15$	0.57	$\pm 0.50$	1.52	$\pm 1.25$	
	Neu%	55.5	$\pm 9.0$	61.7	$\pm 8.0$	68.2	$\pm 8.0$	
	Lym%	35.7	$\pm 9.0$	27.0	$\pm 8.0$	17.5	$\pm 8.0$	
	Mon%	3.8	$\pm 3.8$	3.8	$\pm 3.8$	5.5	$\pm 4.5$	
	Eos%	5.0	$\pm 4.0$	7.5	$\pm 6.0$	8.8	$\pm 7.0$	
	RBC $\times 10^{12}/L$	2.04	$\pm 0.18$	4.44	$\pm 0.24$	5.05	$\pm 0.30$	
	HGB g/L	56	$\pm 4$	130	$\pm 6$	157	$\pm 8$	
	HCT %	18.4	$\pm 1.5$	43.7	$\pm 2.0$	53.4	$\pm 2.4$	
	MCV fL	90.0	$\pm 5.0$	98.5	$\pm 5.0$	105.8	$\pm 5.0$	
	MCH pg	27.5	$\pm 2.5$	29.3	$\pm 2.5$	31.1	$\pm 2.5$	
	MCHC g/L	305	$\pm 30$	297	$\pm 30$	294	$\pm 30$	
	RDW-CV %	13.8	$\pm 3.0$	14.2	$\pm 3.0$	13.5	$\pm 3.0$	
	RDW-SD fL	53.8	$\pm 8.0$	59.7	$\pm 8.0$	63.0	$\pm 8.0$	
	PLT $\times 10^9/L$	46	$\pm 20$	212	$\pm 40$	462	$\pm 60$	
	MPV fL	8.4	$\pm 3.0$	9.0	$\pm 3.0$	9.1	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.188	$\pm 0.100$	0.418	$\pm 0.200$	
	PDW*	17.7	$\pm 3.0$	17.6	$\pm 3.0$	16.2	$\pm 3.0$	

\* For Research Use Only

\*\* These parameters are not provided on BC-5000/BC-5000 Vet analyzers

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.



Mindray

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen 518057, P.R.China

Tel: +86 755 81888998

Fax: +86 755 26582680

EC REP

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726